IEC Appliance Inlet C14 with Filter, "Lock and Shield" Mounting



GRF4: with shield and line filter

C14

70° C

Description

- Panel mount :
- Snap-in version rear-side
- 3 Functions:

Appliance Inlet Protection class I , with EMC-shield , Line filter in standard and medical version

See below:

Approvals and Compliances

Characteristics

- Time saving "Lock and Shield" snap-in design for rear-panel mounting
- All single elements are already wired
- Innovative EMC-shielding with flexible steel segments for an excellent contact to the panel
- Suitable for use in medical equipment according to IEC/UL 60601-1 (1 MOOP, 1 MOPP)

For applications according IEC/UL 62368-1 we recommend variants with bleed resistor $\,$

Other versions on request

- Panel thickness 1.2 mm and 2.0 mm

References

Alternative: version without line filter GRF2

Alternative: Standard version

Weblinks

pdf data sheet, html datasheet, General Product Information, Approvals, Distributor-Stock-Check, Accessories, Detailed request for product

| Technical Data | |
|---------------------------------|---|
| Ratings IEC | 0.5 - 10A @ Ta 40 °C / 250 VAC; 50 Hz |
| Ratings UL/CSA | 0.5 - 15 A @ Ta 40 °C / 250 VAC; 60 Hz |
| Leakage Current | standard < 0.5 mA (250 V / 60 Hz) medical < 5 µA (250 V / 60 Hz) |
| Dielectric Strength | > 1.7 kVDC between L-N > 2.7 kVDC between L/N-PE Test voltage (2 sec) |
| Allowable Operation Temperature | -25 °C to 85 °C |
| Climatic Category | 25/085/21 acc. to IEC 60068-1 |
| IP-Protection | front side IP40 acc. to IEC 60529 |
| Protection Class | Suitable for appliances with protection class I acc. to IEC 61140 |
| Terminal | Quick connect terminals 4.8 x 0.8 mm |
| Panel Thickness S | Screw |

Mounting screw torque max 0.5 Nm

Thermoplastic, black, UL 94V-0

Snap-in: 1.5 mm (other on request)

| Appliance inlet/-outlet | C14 acc. to IEC 60320-1, UL 60320-1, CSA C22.2 no. 60320-1 (for cold conditions) pin-temperature 70 °C, 10A, Protection Class I |
|-------------------------|--|
| Line Filter | Standard and Medical Version, IEC 60939, UL 1283, CSA C22.2 no. 8 Technical Details |
| MTBF | > 3'300'000 h acc. to MIL-HB-217 F |
| | |

Approvals and Compliances

Material

Detailed information on product approvals, code requirements, usage instructions and detailed test conditions can be looked up in Details about Approvals

SCHURTER products are designed for use in industrial environments. They have approvals from independent testing bodies according to national and international standards. Products with specific characteristics and requirements such as required in the automotive sector according to IATF 16949, medical technology according to ISO 13485 or in the aerospace industry can be offered exclusively with customer-specific, individual agreements by SCHURTER.

Approvals

The approval mark is used by the testing authorities to certify compliance with the safety requirements placed on electronic products. Approval Reference Type: GRF4

| Approval Logo | Certificates | Certification Body | Description |
|------------------------------|---------------|--------------------|--------------------------------|
| ₽ | VDE Approvals | VDE | VDE Certificate Number: 102348 |
| c FU ° _{IIS} | UL Approvals | UL | UR File Number: E72928 |

Product standards

Product standards that are referenced

| Organization | Design | Standard | Description |
|--------------|-----------------------|-----------------------|---|
| <u>IEC</u> | Designed according to | IEC 60320-1 | Appliance couplers for household and similar general purposes |
| <u>IEC</u> | Designed according to | IEC 60939 | Passive filters for suppressing electromagnetic interference |
| <u>IEC</u> | Designed according to | IEC 61058-1 | Switches for appliances. Part 1. General requirements |
| (UL) | Designed according to | UL 60320-1 | Standard for Attachment Plugs and Receptacles |
| (UL) | Designed according to | UL 1283 | Passive filters for suppressing electromagnetic interference |
| GSA Group | Designed according to | CSA C22.2 no. 60320-1 | General Use Receptacles, Attachment Plugs, and Similar Wiring Devices |
| CSA Group | Designed according to | CSA C22.2 no. 8 | Electromagnetic interference (EMI) filters |

Application standards

Application standards where the product can be used

| Organization | Design | Standard | Description |
|--------------|--------------------------------|----------------|--|
| <u>IEC</u> | Suitable for applications acc. | IEC/UL 62368-1 | Audio/video, information and communication technology equipment - Part 1: Safety requirements |
| <u>IEC</u> . | Suitable for applications acc. | IEC 60601-1 | Medical electrical equipment - Part 1: General requirements for basic safety and essential performance |


Compliances


The product complies with following Guide Lines


| Identification | Details | Initiator | Description |
|----------------|--------------------------------|-------------|---|
| C€ | CE declaration of conformity | SCHURTER AG | The CE marking declares that the product complies with the applicable requirements laid down in the harmonisation of Community legislation on its affixing in accordance with EU Regulation 765/2008. |
| UK CA | UKCA declaration of conformity | SCHURTER AG | The UKCA marking declares that the product complies with the applicable requirements laid down in the British Amendment of Regulation (EC) 765/2008. |
| ROHS | RoHS | SCHURTER AG | Directive RoHS 2011/65/EU, Amendment (EU) 2015/863 |
| © | China RoHS | SCHURTER AG | The law SJ / T 11363-2006 (China RoHS) has been in force since 1 March 2007. It is similar to the EU directive RoHS. |
| REACH | REACH | SCHURTER AG | On 1 June 2007, Regulation (EC) No 1907/2006 on the Registration, Evaluation, Authorization and Restriction of Chemicals 1 (abbreviated as "REACH") entered into force. |
| 300 | Medical Equipment | SCHURTER AG | Suitable for use in medical equipment according to IEC/UL 60601-1 (1 MOOP, 1 MOPP) |

Dimension [mm]

GRF4








This area must be conductive for optimal shielding. Do not apply paint or coatings.


Assembly Instructions

Step 1



Place the socket in top edge of the panel cut-out.

Step 3




Step 2




Push the socket upwards.

Step 4



Release the socket. The component is self adjusting.



Push the socket forward into the panel cut-out.


Technical Data of Filter-Components


| Rated Current [A] | Filter-Type | Inductances L [mH] | Capacitance CX [nF] | Capacitance CY [nF] | R [M Ω] |
|----------------------|----------------------|-----------------------|------------------------|------------------------|------------------------|
| 0.5 | Standard version | 2 x 24 | 100 | 2.2 | - |
| 1 | Standard version | 2 x 12 | 100 | 2.2 | - |
| 3 | Standard version | 2 x 2.5 | 100 | 2.2 | - |
| 6 | Standard version | 2 x 0.78 | 100 | 2.2 | - |
| 10 | Standard version | 2 x 0.225 | 100 | 2.2 | - |
| 15 | Standard version | 2 x 0.1 | 100 | 2.2 | - |
| 0.5 | Medical Version (M5) | 2 x 24 | 100 | - | 1 |
| 1 | Medical Version (M5) | 2 x 12 | 100 | - | 1 |
| 3 | Medical Version (M5) | 2 x 2.5 | 100 | - | 1 |
| 6 | Medical Version (M5) | 2 x 0.78 | 100 | - | 1 |
| 10 | Medical Version (M5) | 2 x 0.225 | 100 | - | 1 |
| 15 | Medical Version (M5) | 2 x 0.1 | 100 | - | 1 |

Diagrams

Standard version

Medical Version (M5)





- - - - 50Ω differential mode _


1) Line 2) Load 1) Line 2) Load

Attenuation Loss


Standard version

Stariuaru versior


0.5 A




1 A



3 A




6 A




 50Ω common mode

10 A




15A




Medical version (M5)


0.5 A




1 A




3 A




6 A



10 A



15A



All Variants

| Rated Current IEC [A] | Rated Current UL [A] | Filter-Type | Panel Thickness s [mm] | Packaging unit | Order Number |
|--------------------------|-------------------------|----------------------|---------------------------|----------------|-----------------|
| 0.5 | 0.5 | Standard version | 1 | 20 | GRF4.0411.011.C |
| 1 | 1 | Standard version | 1 | 20 | GRF4.0412.011.C |
| 6 | 6 | Standard version | 1 | 20 | GRF4.0416.011.C |
| 10 | 10 | Standard version | 1 | 20 | GRF4.0417.011.C |
| 0.5 | 0.5 | Standard version | 1.5 | 20 | GRF4.0411.013.C |
| 1 | 1 | Standard version | 1.5 | 20 | GRF4.0412.013.C |
| 3 | 3 | Standard version | 1.5 | 20 | GRF4.0413.013.C |
| 6 | 6 | Standard version | 1.5 | 20 | GRF4.0416.013.C |
| 10 | 10 | Standard version | 1.5 | 20 | GRF4.0417.013.C |
| 15 | 15 | Standard version | 1.5 | 20 | GRF4.0419.013.C |
| 3 | 3 | Standard version | 2 | 20 | GRF4.0413.014.C |
| 0.5 | 0.5 | Medical Version (M5) | 1.5 | 20 | GRF4.0021.013.C |
| 1 | 1 | Medical Version (M5) | 1.5 | 20 | GRF4.0022.013.C |
| 3 | 3 | Medical Version (M5) | 1.5 | 20 | GRF4.0023.013.C |
| 6 | 6 | Medical Version (M5) | 1.5 | 20 | GRF4.0026.013.C |
| 10 | 10 | Medical Version (M5) | 1.5 | 20 | GRF4.0027.013.C |
| 15 | 15 | Medical Version (M5) | 1.5 | 20 | GRF4.0029.013.C |

Availability for all products can be searched real-time: https://www.schurter.com/en/info-center/support-tools/stock-che distributors

Packaging unit

20 Pcs

Mating Outlets/Connectors

Category / Description



Appliance Outlet Overview complete

| 4787, Mounting: Screw-on mounting, Appliance Outlet: IEC Solder terminals, 10 A, Suitable for appliances with protection class I | 4787 |
|--|------|
| IEC Appliance Outlet F or H, Screw-on Mounting, Front Side, Solder, PCB or Quick-connect Terminal | 5091 |





| 4782 Mounting: Power Cord, 3 x 1 mm 2 / 3 x 18 AWG, Cable, Connector: IEC C13 | 4782 |
|---|---------|
| 4785 Mounting: Power Cord, 3 x 1 mm² / 3 x 18 AWG, Cable, Connector: IEC C13 | 4785 |
| 4300-06 Mounting: Power Cord, 3 x 1 mm ² / 3 x 18 AWG, Cable, Connector: IEC C13 | 4300-06 |
| 4781 Mounting: Power Cord, Cable, Connector: IEC C15 | 4781 |
| 4784 Mounting: Power Cord, 3 x 1 mm² / 3 x 18 AWG, Cable, Connector: IEC C15 | 4784 |